

You can use a sapphire rod or glass capillary to connect the sample holder and the sample.

Due to the low thermal conductance of the glass, glass capillary is not recommended for precise temperature control